

QUANTITATIVE DETERMINATION OF PRIMARY AROMATIC AMINES IN RECYCLED COLD-CURE AND FLEXIBLE FOAMS

Juliane Böttcher, Kate Monks; applications@knauer.net
KNAUER Wissenschaftliche Geräte GmbH, Hegauer Weg 38, 14163 Berlin; www.knauer.net

SUMMARY

A reliable method for the quantification of aryl amines in cold-cure and flexible foams from recycled mattresses is described in the following application. The focus was set on two substances, 2,4-diaminotoluene (TDA) and 4,4-diaminodiphenylmethane (DAPM), which are mandatory to be determined before processing the recycled foam due to their carcinogenic properties.

INTRODUCTION

Cold-cure and flexible foams used for the production of mattresses are made of polyurethane synthetics. During the synthesis of urethanes, which are the basis for polyurethanes, different intermediate products are formed.

Two of these products occurring in the process are DAPM and TDA. Both substances are classified as carcinogenic. Furthermore TDA is presumed to be teratogenic and mutagenic. When recycling the cold-cured and flexible foams it is necessary to determine the concentration of these compounds before reusing the foams. Referring to OEKO-TEX® Standard 100 a limit value of 20 mg/kg for aryl amines is appointed [1].

RESULTS

For the quantification a calibration was made. Therefore a mixed standard of DAPM and TDA at five different concentrations was used. For both components a correlation coefficient of $R^2=0.999$ was achieved. Exemplary one sample of flexible foam was selected and spiked with standard to a concentration of 0.1 mg/mL. **Fig. 1** shows the sample measurement and **Fig. 2** shows an overlay of the sample (red) and spiked sample (blue) of flexible foam. The limit of detection (LOD) was determined with 0.63 µg/mL for TDA and 0.67 µg/mL for DAPM.




Fig. 1 Sample of flexible foam (18 mg/mL)




Fig. 2 Overlay of sample (red) and spiked sample (blue) of flexible foam

MATERIALS AND METHOD

An AZURA Analytical HPLC Plus system for a pressure range up to 700 bar was used for this application. It consisted of an AZURA P 6.1L LPG pump, an autosampler 3950, an AZURA CT 2.1 column thermostat and an AZURA MWD 2.1L multiwavelength detector. The analytical method was run with a step gradient at a flow rate of 1.0 mL/min. The mobile phase was a mixture of water and acetonitrile, both with 0.1% triethylamine as mobile phase modifier. The column thermostat was set to 25 °C and the detector recorded at 290 nm. The column that was used was filled with ProntoSIL 120-3 C8 ace EPS silica.

CONCLUSION

With the developed method and the AZURA HPLC Plus system it was possible to perform a rapid quantitative analysis of 2,4-diaminotoluene and 4,4-diaminodiphenylmethane without time consuming sample preparation. Even a complex matrix such as the recycled cold-cured and flexible foams can be determined robust and reproducible with the specified method parameters.

REFERENCES

[1] https://www.oeko-tex.com/de/business/certifications_and_services/ots_100/ots_100_limit_values/ots_100_limit_values.html

Additional information:



QUANTITATIVE DETERMINATION OF PRIMARY AROMATIC AMINES IN RECYCLED COLD-CURE AND FLEXIBLE FOAMS

ADDITIONAL RESULTS




Fig. A1 Mixed standard of TDA and DAPM (both 0.1mg/mL)

ADDITIONAL MATERIALS AND METHODS

Tab. A1 Method parameters

Eluent A	H ₂ O _{dd} +0.1 % TEA		
Eluent B	Acetonitrile+0.1 % TEA		
Gradient	Time [min]	% A	% B
	0	95	5
	5	95	5
	15	35	65
	15.02	0	100
	25	0	100
	25.02	95	5
	35	95	5
Flow rate	1 mL/min	System pressure	ca. 190 bar
Column temperature	25 °C	Run time	35 min
Injection volume	10 µL	Injection mode	Full loop
Detection wavelength	290 nm	Data rate	20 Hz
		Time constant	0.05 sec

Tab. A2 System configuration & data

Instrument	Description	Article No.
Pump	AZURA P 6.1L LPG, 10 ml, SSt	APH35EA
Autosampler	Autosampler 3950	A50070
UV Detector	AZURA MWD 2.1L	ADB01
Flow cell	LightGuide 10 mm, 2 µL	AMC19
Thermostat	AZURA CT 2.1	A05852
Eluent tray	AZURA E 2.1L	AZC00
Column	Vertex Plus Column 150 x 4.6 mm ProntoSIL 120-3 C8 ace EPS with precolumn	15VF08APSG
Software	OpenLAB CDS EZChrom Edition	A2600-1

